



# INFLUENZA A H3N2

## MULTIPLEX RT-qPCR TEST HANDBOOK

For Research Use Only. Not intended for diagnostic use.

**CAT NO.: YSL-qPX-EC-H3N2-100**

100 reactions  
With Endogenous Control and Lyophilised MasterMix

**VERSION 3.0**



YouSeq Ltd  
8 Moorside Place  
Moorside Road  
Winchester  
SO23 7FX  
United Kingdom

+44 (0) 333 577 6697  
[hello@youseq.com](mailto:hello@youseq.com)

[youseq.com](http://youseq.com)

## INTENDED USE

This RT-qPCR test offers an efficient and user-friendly solution for the detection of Influenza A H3N2 (H3N2) in extracted nucleic acid samples from a variety of sources. It is intended for use by trained professionals in a suitable molecular biology laboratory.

## SPECIFICITY AND SENSITIVITY

### Specificity

The YouSeq RT-qPCR test for detection of Influenza A H3N2 is designed to have the broadest detection profile possible and detect all clinically relevant strains. The primers and probes typically have a  $\geq 95\%$  homology with all reference data used, from relevant, publicly available databases at the time of design.

The target genes outlined below, have been demonstrated to have distinctive sequences making them ideal targets for highly specific detection.

Influenza A H3: HA gene  
Influenza A N2: NA gene

For further information on the detection profile of the product, please do not hesitate to contact our team: [support@youseq.com](mailto:support@youseq.com)

### Sensitivity

The RT-qPCR test is suitable for the detection of these targets across a wide dynamic range. Under ideal PCR conditions the assay can detect less than 100 copies of the targets in the PCR reaction.

## CONTENTS

Component	Cap Colour	Volume
Multiplex primer/probe mix: H3 (FAM) N2 (ROX) Endogenous Control (CY5)		100 µL
Lyophilised Tetra™ OneStep 2X RT-qPCR MasterMix		1.1 mL *
MasterMix Resuspension Buffer (MMRB)		1.5 mL
H3N2 Positive Control (PTC)		500 µL *
Template Resuspension Buffer (TRB)		1.5 mL
DNase/RNase Free Water		1.5 mL

\* Supplied dried – requires resuspension. See instructions in resuspension section.

## RECOMMENDED ADDITIONAL REAGENTS & MATERIALS

Nucleic acid extraction kit

General laboratory equipment (pipettes, pipette tips, (micro)centrifuge tubes, compatible strip tubes/plates, plate seals, etc.)

qPCR instrument with channels to detect FAM, CY5 and ROX.

## BEST PRACTICE

### Decontamination:

Before beginning laboratory work, thoroughly decontaminate any work surfaces and pipettes being used, to eliminate potential contamination.

### General use and set-up:

All components should be fully defrosted with contents at the bottom of the tube before opening. To ensure contents are at the bottom, centrifuge or gently tap the tube. After use, reagents should be returned to the freezer.

Once any reagents are resuspended, mark the tick box on the tube for future reference. After this, or after combining reagents, the tube should be pulse vortexed to ensure it is mixed well.

It is advised to set up the tubes/plate and reaction mix on ice to minimise artefact formation, which may reduce sensitivity.

When preparing the qPCR reaction mix, it is recommended to incorporate an overage when calculating the total number of reactions to compensate for potential volume losses incurred during pipetting.

### Set-up environments:

It is best practice to set up qPCR tubes/plates in two different environments - a clean (no template) lab and PCR (template) lab.

No Template Control(s) (NTC) and Positive Control(s) (PTC) should be included in every run. To reduce contamination, NTCs and samples can be set up and sealed in a clean lab before moving to the PCR lab.

## BENCH SIDE PROTOCOL

### RESUSPENSION

Before first use, resuspend the designated component with the correct reagent and specified volume, as per the table below:

1. Add the resuspension reagent and pulse vortex the tube to ensure each is mixed well.

Component	Reagent	Volume	Location
Lyophilised Tetra™ OneStep 2X RT-qPCR MasterMix	MMRB	1.1 mL	Clean lab
H3N2 Positive Control (PTC)	TRB	500 µL	PCR lab

## RT-qPCR REACTION SET-UP

1. Retrieve the required components and appropriate plasticware for RT-qPCR reaction set-up.
2. In an appropriately sized (micro)centrifuge tube, combine the following reagents to create a reaction mix that will cover all required wells (e.g. **sample(s)**, **NTCs** and **PTCs**).

**Please note:** When calculating required reactions, include an overage to allow for volume loss during pipetting


Component	Volume per reaction
Tetra™ OneStep 2X RT-qPCR MasterMix	10 µL
H3N2 specific primer/probe mix	1 µL
Reaction mix volume	11 µL

**Please note:** It is recommended to work quickly with these components, on ice. The activity of Reverse Transcriptase enzymes at room temperature can create artefacts, which may impact assay sensitivity. YouSeq recommends the tube/plate set-up is no longer than 30 minutes.

3. Mix the combined reagents by briefly vortexing or inverting.
4. Dispense 11 µL of the reaction mix into all required tubes/wells.
5. For the **NTC(s)**, add 9 µL of DNase/RNase Free Water into required tube/well(s).
6. For each extracted **sample**, add 9 µL into required tube/well(s).

**Please note:** It is best practice to seal the **NTC(s)** and **sample** tube/wells before proceeding to the positive control step.

7. For the **positive control(s)**, add 9 µL of the resuspended **PTC**, into the required tube/well(s).
8. Seal the tube(s)/plate.
9. Briefly vortex the sealed tube(s)/plate, then spin in a centrifuge to ensure all reagents are fully resuspended and at the bottom of the tube(s)/wells before proceeding.



## RT-qPCR AMPLIFICATION PROTOCOL

1. Load the tubes/plate onto the qPCR instrument and set up the RT-qPCR protocol following the table below.
2. Set the total reaction volume to 20 µL.

**Please note:** If using a qPCR instrument that uses ROX as a passive reference, ensure the passive reference is turned off or set to "none" before starting the run.

Temperature	Time	Number of Cycles
55°C	10 minutes	-
95°C	3 minutes	-
95°C	15 seconds	x 45
60°C	60 seconds	

Collect fluorogenic data through FAM, CY5 and ROX channels during this step.

3. Start the run.

## INTERPRETATION OF RESULTS – OVERVIEW

If using single threshold analysis – YouSeq recommends setting the threshold at 10% of the End Point Fluorescence (EPF) for each channel.

- For the Target channel, use the **PTC** EPF to set the threshold.
- For the Endogenous Control, use the average EPF from **samples** to set the threshold.

### Results interpretation:

Reaction Type		qPCR Signal (Cq)			
Positive control		18 ± 2	18 ± 2	18 ± 2	18 ± 2
No template control		-	-	-	-
Sample	Endogenous Control: (CY5)	≤ 28	≤ 28	≤ 28	≤ 28
	H3 (FAM)	+	+	-	-
	N2 (ROX)	+	-	+	-
Result		Positive result: Influenza A H3N2	Positive result: Influenza A H3	Positive result: Influenza A N2	Negative result

## INTERPRETATION OF RESULTS – CONTROLS EXPLAINED

### Positive control

The **PTC** should amplify in a Cq range of  $18 \pm 2$  for each target. If this Cq range is not achieved, the sample test result for the associated target is invalid and should be repeated.

**Please note:** The positive control is a sequence representative of the target regions and does not contain the organism's entire genome. The positive control does not include the Endogenous Control sequence and should not be expected to amplify in the CY5 channel.

### No template control

The **NTC** well(s) should be negative, with no amplification.

**Please note:** Background laboratory contamination can result in a very late signal in **NTC** wells. If the **NTC** has amplification, comparison to the sample test result is necessary:




- If the **NTC** is  $\geq 5$  Cq later than the sample signal, it can be considered negative, and the sample test result is valid.
- If the **NTC** is  $< 5$  Cq later, the sample test result is invalid – the test should be repeated after potential sources of contamination have been removed.

### Endogenous Control

Detection of the Endogenous Control is through the CY5 channel. The Cq value from the Endogenous Control will vary according to the amount of RNA in the sample. Cq value of  $\leq 28$  indicates a successful extraction has taken place with sufficient host derived RNA in the sample. If the signal is later than this, repeating the nucleic acid extraction is advised.

**Please note:** If the sample delivers a strong positive result for the target(s) of interest, then amplification of the Endogenous Control may be affected and may appear later. In this case, lack of Endogenous Control amplification is acceptable, and the sample test result is valid.

## MULTIPLEX TROUBLESHOOTING

	1	2	3
Trace			
What can you see?	Different fluorescence levels between Targets	Unexpected low level of fluorescence, similar to amplification from a different channel, within the same well	Amplification for late positives have lower fluorescence, or appear "flatter" than the positive control
Cause	Fluorescence depends on the assay design and the channel e.g. FAM channel is typically the brightest	'Bleed-through' between channels; Amplification from one channel has been incorrectly assigned to a different channel e.g. fluorescence in FAM 'bleeding-through' to CY5	PCR artefact formation during the run (typically driven by Reverse Transcriptase)
Action	Analyse each channel individually so the Y-axis is appropriate for that assay	Ensure instrument is compatible with the dye(s) used in this test and/or recalibrate qPCR instrument	Minimise the time between creating reaction mix (primer/probe + MasterMix) and starting the run. Prepare the reaction mix and plate set-up on ice/cold block.

## PRODUCT SPECIFICATIONS

### Storage

Store at -20°C from arrival. The product's shelf life is outlined as an expiry date on the pouch label.

### Suitable input material

This RT-qPCR test will work well with any source of good quality nucleic acid. Good quality is defined as nucleic acid with high integrity (not degraded). Poor quality input nucleic acid is a leading cause of test failure.

### Regulatory status

This product has been developed for Research Use Only (RUO) and is not intended for diagnostic use. It should not be used for diagnosis of disease or infection unless specifically approved by the regulatory authorities in the country of use.

### Quality Control

In accordance with the YouSeq Ltd ISO EN 13485-certified Quality Management System, each lot of YouSeq Influenza A H3N2 Multiplex RT-qPCR Test is tested against predetermined specifications to ensure consistent product quality. The primers/probe(s) typically demonstrate  $\geq 95\%$  *in silico* specificity to their intended target and periodically checked against newly available sequence information to maintain their detection profile.

### Technical Assistance

For customer support, please contact:

e-mail: [support@youseq.com](mailto:support@youseq.com)  
phone: +44 (0)333 577 6697

### Trademarks and Disclaimers

YouSeq®, Tetra™, ROX™, VIC™.

Registered names, trademarks, etc. used in this document, even if not specifically marked as such, are not to be considered unprotected by law.

Not available in all countries.

© 2025 YouSeq Ltd; all rights reserved.